



LEADING THE WORLD IN PERFORMANCE-BASED VALIDATION

June 16, 2006

Dr. Morgan Wallace
DuPont Qualicon
ESL Building 400
PO Box 80400
Route 141 & Henry Clay Road
Wilmington, Delaware 19880-0400

RE: Approval of a Level 2 Modification of the BAX® System PCR Assay for Screening *Listeria* genus using the BAX® Q7 Instrument.

Dear Dr. Wallace:

Congratulations! The Level 2 Modification of the BAX® System PCR Assay for Screening *Listeria* genus using the BAX® Q7 Instrument has been approved for *Performance Tested Methods*SM status. This modification validates the use of the BAX® Q7 instrument in addition to the originally validated BAX® instrument for use with the BAX® System PCR Assay for Screening *Listeria* genus. The data from this evaluation demonstrated equivalent performance of the newly modified method with the performance of the USDA cultural method. This kit is assigned certification license number 030502.

Kind regards,

Maria Nelson
Technical Consultant
AOAC Research Institute
AOAC INTERNATIONAL

☞ CONTACT US ☞

📍 481 N. FREDERICK AVE, SUITE 500, GAITHERSBURG, MD 20877-2417 ☎ +1 (301) 924-7077 📠 +1 (301) 924-7089

🌐 [HTTP://WWW.AOAC.ORG/TESTKITS/TESTKITS.HTML](http://www.aoac.org/testkits/testkits.html)